

Certificate of Analysis

TAKARA

Product name/description : GT-T551 Culture medium, 1L Bottle

Cat. # : SK551S
Lot. # : 0520150D
Storage Condition : 4 degrees C
Shipping Condition : 4 degrees C
Expiration Date : April 30, 2016
Package Size: 1,000 ml

Quality Control Data :

Test items	Specifications	Results
1) Sterility Testing	Negative	Negative
2) Mycoplasma Testing	Negative	Negative
3) pH	7.2~7.6	7.4
4) Endotoxin Testing	≤ 0.25 EU/ml	0.004 EU/ml
5) Osmolarity	300 ± 10 mOsm/kg	300 mOsm/kg
6) Cell Culture Testing	Passed	Passed

It is certified that this product meets above specifications.

Approved By :

Manager, Quality Assurance
TAKARA BIO INC.

Safety Information :

Please refer to our website for safety information :

<http://www.takara-bio.com>

Notice To Purchaser :

This product is not intended for human in vivo applications. It is the end user's responsibility to ensure that the final product meets the requirements of the application for which it is to be used.

This product may not be resold or transferred, modified for resale or transfer without written approval from TAKARA BIO INC.

If you require licenses for other use, please contact us by phone at +81 77 543 7247 or from our website.

<http://www.takara-bio.com>

Your use of this product is also subject to compliance with any applicable licensing requirements described on the product web page.

It is your responsibility to review, understand and adhere to any restrictions imposed by such statements.

All trademarks are the property of their respective owners. Certain trademarks may not be registered in all jurisdictions.

Country of origin: Japan